

SOLUTION CHANGE

 WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 11/01/2021

OPERATOR NAME:
 TREVOR J PECK
 Permit #: 49747
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:14
INTERNAL STANDARD	VERIFIED	06:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:16
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:18

78P #637

