

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/16/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:30
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:33
BLANK TEST	0.00	02:34

