

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 12/13/2021

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:52
INTERNAL STANDARD	VERIFIED	20:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:52
BLANK TEST	0.00	20:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:54
BLANK TEST	0.00	20:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56

